Oxygen Sensor OOM105

ENVITEC

Measurement Range: 0-100 % oxygen

Output in ambient air: Compatible with Teledyne® T-7 sensor

Electrical Interface: Molex plug 4P4C

Accuracy and Repeatability: < 1 % vol. O2 when calibrated at 100 %

Oxygen

Linearity error: < 3 % relative

Response time: < 5sec. to 90 % of final value **Zero Offset Voltage:** < 100 μ V in 100 % nitrogen

applied after 5 min

Cross Interference: < 0.5 % vol. O₂ response to:

10 % CO₂ balance N₂ 80% N₂O balance N₂

7.5% Halothane balance N_2 7.5% Isoflurane balance N_2 7.5% Enflurane balance N_2 9% Sevoflurane balance N_2 20% Desflurane balance N_2

Influence of Humidity: 20% Desilurane balance N₂ - 0.03 % rel. per % RH at 25°C

Influence of Pressure: proportional to change in oxygen partial

oressure

Influence of Mechanical Shock: < 1% relative after a fall from 1m

Operating Temperature: 0 to 50°C

Temperature Compensation: built-in NTC compensation

Effect of Temperature between +25 °C and +40 °C: 3 % relative

Compensation (steady state): erro

between 0 °C and +50 °C: 8 % relative error

Operating Humidity: 0-99 % RH non-condensing **Long Term Output Drift:** < 1 % vol oxygen per month

typically < - 15 % relative over lifetime


Storage Temperature: $-20 \text{ to } +50 \text{ }^{\circ}\text{C}$ **Recommended Storage:** $+5 \text{ to } +15 \text{ }^{\circ}\text{C}$

Warm-Up Time: < 30 minutes, after replacement of sensor

Nominal Sensor Lifetime: ≥ 500.000 % vol oxygen hours
Weight: approximately 28 grams

Warranty Period: 15 months
Part No.: 01-00-0053

All specifications are applicable at standard conditions: 1013 hPa, 25°C dry ambient air



A: NO CONNECTION

B: (+)

Dimension in mm

Use the advantages:

- Meet EN ISO 21647
- Designed and manufactured according to EN ISO 9001 : 2000 and EN ISO 13485 : 2003
- Accurate and reliable fast response
- Resistant to N₂O
- Excellent signal stability
- High product quality
- Short lead times
- Technical support



ENVITEC- WISMAR GMBH TEL:13361073697(China)

Alter Holzhafen 18 D-23966 Wismar

Phone: +49-(0) 3841- 360 1 Fax: +49-(0) 3841- 360 222 E-Mail: <u>info@envitec.com</u> http://www.envitec.com