Linearity error:

Oxygen Sensor OOM110

ENVITE

Measurement Range: 0-100 % oxygen **Output in ambient air:** 9 to 12mV **Electrical Interface:** Modular Jack 6P4C

Accuracy and Repeatability: < 1 % vol. O2 when calibrated at 100 %

> Oxygen < 3 % relative

Response time: < 12sec. to 90 % of final value **Zero Offset Voltage:** < 200 μV in 100 % nitrogen

applied after 5 min

Cross Interference: < 0.5 % vol. O_2 response to:

> 10 % CO₂ balance N₂ 80% N₂O balance N₂

7.5% Halothane balance N₂ 7.5 % Isoflurane balance N₂ 7.5 % Enflurane balance N₂ 9% Sevoflurane balance N₂ 20% Desflurane balance N₂

Influence of Humidity: - 0.03 % rel. per % RH at 25°C **Influence of Pressure:** proportional to change in oxygen partial

Influence of Mechanical Shock: < 1% relative after a fall from 1m

Operating Temperature: 0 to 50°C

Temperature Compensation: built-in NTC compensation

Effect of Temperature between +25 °C and +40 °C: 3 % relative

Compensation (steady state):

between 0 °C and +50 °C: 8 % relative error

Operating Humidity: 0-99% RH non-condensing **Long Term Output Drift:** < 1 % vol oxygen per month

typically < - 15 % relative over lifetime

Storage Temperature: -20 to +50 °C **Recommended Storage:** +5 to +15 °C **Recommended Load:** ≥ 10 kOhms

Warm-Up Time: < 30 minutes, after replacement of sensor

Nominal Sensor Lifetime: ≥ 1.000.000 % vol oxygen hours Weight: approximately 50 grams

Warranty Period: 15 months


Part No.: 01-00-0098


All specifications are applicable at standard conditions: 1013 hPa, 25°C dry ambient air

Use the advantages:

- Meet EN ISO 21647
- Designed and manufactured according to EN ISO 9001: 2000 and EN ISO 13485: 2003
- Accurate and reliable response
- Resistant to N₂O
- Excellent signal stability
- High product quality
- Short lead times
- Technical support







Dimension in mm

ENVITEC-WISMAR GMBH TEL:13361073697(China)

Alter Holzhafen 18 D-23966 Wismar

Phone: +49-(0) 3841- 360 1 Fax: +49-(0) 3841- 360 222 E-Mail: info@envitec.com http://www.envitec.com